



EVA Select Digital Sensor Care Do's and Don'ts

In an effort to help you keep your EVA Select digital sensor in good working condition, please follow these simple Do's and Don'ts. Proper sensor care is easy and damage can often be avoided by following these tips.

Do:

- DO wipe down the sensor using a FEW drops of disinfecting solution on a cotton swab
- DO store the sensor in its docking station when it is not in use
- DO use proper bite holders, tabs and sterile sheaths every time the sensor is used
- DO always uncoil sensor gently
- DO carefully remove all tangles and sharp bends while uncoiling cable
- DO ensure cable is completely uncoiled and has no kinks or sharp bends in it before each use
- DO grasp the connector, not the cable, when disconnecting the sensor from the docking station, and pull gently
- DO hold onto the sensor body while removing any tabs, holders, or sheaths. Be as gentle as possible

Don't:

- **DON'T use ANY liquids to wipe down the sensor, including disinfecting wipes.**
- DON'T autoclave the sensor
- DON'T soak or directly spray the sensor or its connector with disinfecting solution
- DON'T clamp the sensor or its cable with a hemostat or equivalent
- DON'T pull the sensor by its cable during sheath removal
- DON'T hang the sensor by its cable
- DON'T allow anyone to bite on the sensor
- DON'T let the cable hang on or near the floor where it can become tangled
- DON'T let the cable hang near a drawer where it can be kinked/crimped
- DON'T pull on the cable to remove from tab, holder or sheath
- DON'T tangle cable during use
- DON'T create sharp bends and/or knots when untangling

Please Note: EVA Select sensors are not water-proof, and should not be treated as liquid-resistant.